



Second Nature **Security**

We declare that the PulseOn Arrhythmia Monitor has undergone and passed rigorous penetration testing. The testing was conducted by an experienced OffSec Certified Professional (OSCP). Our standard testing process strictly adheres to the vulnerabilities outlined in the OWASP Top 10, leveraging industry-standard methodologies, including the Open Source Security Testing Methodology Manual, OWASP Testing Guide and OWASP Application Security Verification Standard (ASVS). The following areas were the main focus of the test:

- Configuration and deployment
- Identity management
- Authentication
- Authorization
- Session management
- Input validation
- Error handling
- Cryptographic security
- Business logic controls
- Denial of service
- Client-side security
- API-level security

The test results demonstrated that the product performed exceptionally well. All uncovered vulnerabilities were below the Common Vulnerability Scoring System (CVSS) score of 7.0. No critical or high-risk vulnerabilities were identified, and all medium- and low-level issues were promptly fixed. The fixes were further verified in a second follow-up testing round.

Juho Ranta, CTO
Second Nature Security
12.8.2024